

APSIC CSSD Center of Excellence Program

Audit Checklist



Please make sure to answer
ALL questions to obtain a
more accurate audit.

Highlighted questions
are **MANDATORY** by
APSIC GUIDELINES





Audit Checklist

Table of Contents

SECTION A: Handling, Collection and Transport of Contaminated Instruments

Question 1:

Reusable items separated from waste at point of use

Not separated
Partial separation
Clear separation

Question 2:

Contaminated disposable items are discarded appropriately (including sharps)

Not discarded appropriately
Partial compliance
Discarded appropriately all the time

Question 3:

Gross soil is removed from instruments at point of use if immediate transportation not possible

Not removed
Removed sometimes
Removed all the time

Question 4:

Soiled items should be kept moist (moist towel, enzyme foam or spray product)

Items are dry
Partially moist
Moist all the time

Question 5:

Secured, dedicated containers are provided for soiled instruments

Not available
Sometimes
All the time

Question 6:

Use of puncture resistant, leak-proof containers for soiled items

Not available
Sometimes
All the time

Question 7:

Soiled items must be contained during transportation

Not evident
Attempts made
Done all the time

SECTION A: Handling, Collection and Transport of Contaminated Instruments (continued)



Question 8:

Transportation of soiled instruments avoids high (public) traffic areas

Not evident

Attempts made

Done all the time

Question 9:

Transportation carts should be covered and should prevent items from falling over or off

Not done

Occasionally done

Done all the time

Question 10:

Dedicated elevators (or lifts) with direct access to decontamination area

Not available

Dedicated but not used as planned all the time

Dedicated and used as planned all the time

Question 11:

Policy and procedure in place for transportation of contaminated items between buildings, if applicable

Not available

Policy exists but not followed

Policy exists and complied with all the time

**SECTION A:
Handling,
Collection and
Transport of
Contaminated
Instruments**

TOTAL:

/ 160 points

MANDATORY ITEMS MARKED:

/ 100 (5 items)

I completed all questions in this section especially the **MANDATORY** (highlighted) items.

Yes

No

SECTION B: Cleaning and Decontamination Processes



Question 1:

Written policies and procedures in place for all cleaning and decontamination processes

Both not available

One is available

Both are available

Question 2:

Instrumentation is disassembled (according to manufacturer's instructions) to expose all surfaces for cleaning

Not done

Done occasionally

Done all the time

Question 3:

Rigid container systems disassembled according to manufacturer instructions (filters, valves and interior baskets)

Not done

Done occasionally

Done all the time

Question 4:

Cleaning agents are used according to manufacturer's instructions (dilution and temperature, etc.)

Not done

Done occasionally

Done all the time

Question 5:

Appropriate manual and mechanical cleaning methods are used according to manufacturer's instructions and IFU's are accessible to decontamination staff

Not done

Done occasionally

Done all the time

Question 6:

Appropriate personal protective equipment (PPE) are used

Not evident

Occasionally practiced

Done all the time

Question 7:

Appropriate brushes/cleaning implements designed for use on medical devices are used

Not done

Done occasionally

Done all the time

SECTION B: Cleaning and Decontamination Processes (continued)



Question 8:

Brushes/cleaning implements are either disposable or if reusable, are decontaminated at least daily

Not evident

Decontaminated irregularly

Decontaminated at least daily

Question 9:

Monitoring of mechanical cleaning equipment should be done upon installation and then weekly (preferably daily) and recorded

Not done

Done on installation but not monitored subsequently

Done on installation and monitored with documentation weekly/daily

Question 10:

Appropriate manual and mechanical rinsing methods are understood and are done according to manufacturer's instructions

Staffs are unaware of these

Some understood and complied

All understood and complied

Question 11:

Cleaning agent (enzymatic cleaner) should be compatible with the medical device to be cleaned

No evidence of compatibility

Some are compatible

All are compatible

Question 12:

Chemical for disinfectants and terminal sterilisation are used according to manufacturer's instructions

Not done

Partial compliance

Full compliance

Question 13:

Ultrasonic cleaner solution is changed at specified frequency or sooner if needed

No frequency set

Frequency set but not complied

Frequency set and complied

SECTION B: CLEANING AND DECONTAMINATION PROCESSES continues >>>

SECTION B: Cleaning and Decontamination Processes (continued)



Question 14:

Final rinse in washer disinfectant is done with treated water (deionized, distilled, or RO water)

Final rinse with untreated water

Final rinse with treated water

**SECTION B:
Cleaning and
Decontamination
Processes**

TOTAL:

/ 180 points

MANDATORY ITEMS MARKED:

/ 80 (4 items)

I completed all questions in this section especially the **MANDATORY** (highlighted) items.

Yes

No

SECTION C: Instrumentation Inspection, Preparation and Packaging



Instrument inspection

Question 1:

Ensure instruments are cleaned and dried before packaging

Not evident

Partial compliance

Full compliance

Question 2:

Inspect instruments for flaws or damage. Check for rust, pitting, corrosion, burrs, nicks, cracks, chipping of plated surfaces. Lighted magnifying glass available for instrument inspection.

Lighted magnifying glass not available

Occasional inspection with lighted magnifying glass done

Inspection done with lighted magnifying glass for all instruments

Question 3:

Cleaning verification by users should include visual inspection combined with other verification methods (ATP) that allow assessment of instrument surfaces and channel

Only visual inspection

Other verification methods included but without fixed schedule and plan

Other verification method included with fixed schedule and plan

Question 4:

Instruments:

- Cutting edges are sharp
- Moving parts move freely, without sticking
- Instruments needing repair are taken out of service for repair or replacement

Not evident

Done occasionally

Done all the time

Question 5:

Follow MDMs instructions for instruments requiring lubrication after cleaning or prior to sterilisation

Not evident

Done occasionally

Done all the time

SECTION C:

Instrumentation Inspection, Preparation and Packaging (continued)



Preparation and assembly

Question 6:

Delicate/sharp instruments are protected while being handled/assembled for sterilisation (*may use special holders, tip guards, or foam sleeves*)
– Tip protectors should be sterilant-permeable

Not evident
Done occasionally
Done all the time

Question 7:

Instruments that open (e.g. scissors, haemostats) are held in unlocked, open positions

Not evident
Done occasionally
Done all the time

Question 8:

Multi-part instruments are disassembled prior to sterilisation, ensuring all parts are easily accessed for aseptic assembly

Not evident
Done occasionally
Done all the time

Question 9:

Lumened devices:
– Remove stylets/plugs, such as catheters, needles, tubings
– Moistening of the lumen may be recommended; consult device manufacturer

Not evident
Done occasionally
Done all the time

Question 10:

Complex instruments (eg, air-powered, endoscopes, having lumens or channels) are prepared according to written IFU from device manufacturer

Not evident
Done occasionally
Done all the time

Question 11:

Non-linting absorbent material may be placed in trays to help facilitate drying. Tray liners or other absorbent materials may be used to alleviate drying problems.

Not evident
Done occasionally
Done all the time

SECTION C:

Instrumentation Inspection, Preparation and Packaging (continued)



Preparation and assembly (continued)

Question 12:

Basins:

- Graduated basins should differ in diameter by one inch
- Use non-linting absorbent material between nested basins
- Wrapped basin sets should not exceed 3kg (7 lbs)

Not evident

Done occasionally

Done all the time

Question 13:

Containerized instrument sets do not exceed 11kg (25 lbs)

Not evident

Occasionally comply

Full compliance all the time

Packaging

Question 14:

Packaging materials are held for a min. of 2 hrs. prior to use at room temp (21°F-24°F) and at a relative humidity ranging from 30-60%. *[This is needed to permit steam sterilisation and prevent superheating.]*

Not evident

One factor complied with

Both factors complied with

Question 15:

Packaging materials are examined regularly for defects (i.e. holes, worn spots, stains)

Not evident

Done occasionally

Done all the time

Question 16:

Wrappers should be kept snug, but not wrapped too tightly or strike-through could occur

Not evident

Done occasionally

Done all the time

Question 17:

Paper/plastic pouches:

- Labeling is done on plastic side only
- Double peel pouch only if pouch is validated for this use

No labeling

Labeling but not on plastic side

Labeling on plastic side

SECTION C: Instrumentation Inspection, Preparation and Packaging (continued)



Packaging (continued)

Question 18:

Wrapped packs:

– Write only on indicator tape or affixed labels

No labeling

Labeling done inappropriately

Labeling on indicator tape or
affixed labels

Question 19:

Perforated, wire-mesh-bottom trays, and rigid
organizing trays are inspected prior to each use
to ensure there are no sharp edges, nicks, or
loose wire-mesh

Not evident

Done occasionally

Done all the time

Question 20:

Tape (other than sterilisation indicator tape)
should not be used to secure packages, nor
should safety pins, ropes, paper clips, staples,
or other sharp objects

Not evident

Done occasionally

Done all the time

Question 21:

Validation test to be done for heat sealer
at set frequency

Not done

Frequency set but not followed

Frequency set and followed

**SECTION C:
Instrumentation
Inspection,
Preparation
and Packaging**

TOTAL:

/ 260 points

MANDATORY ITEMS MARKED:

/ 100 (5 items)

I completed all questions in this section
especially the **MANDATORY** (highlighted) items.

Yes

No

SECTION D: Sterilisation and Monitoring



Follow manufacturers' guidelines

Question 1:

Steriliser manufacturers:

- Written instructions for cycle parameters are available

Not available

Some are available

Available for all sterilisers

Question 2:

Rigid container manufacturers:

- Instructions for cycle parameters are followed

Not evident

Done occasionally

Done all the time

Question 3:

Medical device manufacturers:

- Written instructions for sterilisation cycle parameters are available/accessible for items to be sterilised, including loaner sets

Not available

Available but not easily accessible

Available and easily accessible

Loading the steriliser (follow steriliser manufacturers' written instructions)

Question 4:

Group together similar items requiring same cycle parameters

Not evident

Done occasionally

Done all the time

Question 5:

Steriliser cart:

- Allow space between packs
- Do not overload
- Packages should not touch chamber walls

Not evident

Overloading seen occasionally

Complied with all the time

SECTION D: STERILISATION AND MONITORING: Loading the steriliser continues >>>

SECTION D: Sterilisation and Monitoring (continued)



Loading the sterilizer (continued)

Question 6:

Mixed loads:

- Place metal items on the loading cart below textiles and paper-plastic pouches (to prevent condensate from dripping onto lower packs)

Not evident

Done occasionally

Done all the time

Question 7:

Solid-bottom pans, bowls, and trays are tilted on edge and oriented in the same direction

Not evident

Done occasionally

Done all the time

Question 8:

Paper-plastic pouches:

- Use baskets to facilitate placing pouches on edge

Not evident

Done occasionally

Done all the time

Question 9:

Rigid containers:

- Stacking could interfere with air evacuation; follow container's manufacturer's instructions

Not evident

Done occasionally

Done all the time

Unloading the steriliser

Question 10:

Open steriliser door properly

- Door may be opened slightly at the end of the cycle (for some time) prior to removing the load

Not evident

Done occasionally

Done all the time

Question 11:

Load contents:

- There should be no visible signs of liquid, or water droplets. (Wet items are considered contaminated even if not touched.)

Not evident

Wetness seen occasionally

Dry for all items

SECTION D: Sterilisation and Monitoring (continued)



Unloading the sterilizer (continued)

Question 12:

Sterilised items remain on the cart to cool for a minimum of 30 minutes, and are not touched during the cooling process

Not evident

Done occasionally

Done all the time

Question 13:

Place cart in a low traffic area without proximity to air-conditioning or cold-air vents

Carts placed in close proximity to vents

Some carts are placed in close proximity to vents

Complied with always

Question 14:

Immediate use "Flash" items:

- Are used immediately and not stored for later use (assume condensate will be present)

Not used immediately

Occasionally used immediately

Used immediately always

Physical monitors, chemical indicators and biological indicators

Question 15:

Verify parameters of the cycle have been met by reviewing cycle printout tapes.
Circle minimum temperature and exposure time, initial/sign, and date.

Not evident

Reviewed but not documented regularly

Reviewed and documented all the time

Question 16:

Bowie-Dick Testing is done daily in pre-vacuum sterilisers before first processed load.
Process Bowie-Dick at 132°-134°C for 3.5 to 4 minutes. One pack per load in an empty chamber. Record results.

Not evident

Done occasionally

Done daily

Question 17:

External process indicators (indicator tape, labels) are affixed to hospital-sterilised packages and containers

Not evident

Affixed on some

Affixed on all items

SECTION D: Sterilisation and Monitoring (continued)



Physical monitors, chemical indicators and biological indicators (continued)

Question 18:

Internal chemical indicator(s) (Type 4, 5, 6) are placed inside every package in the most challenging location for sterilant to reach (refer to Rigid Container Manufacturers' Instructions for CI placement)

Not evident

Placed in some items/not in most challenging location

Complied with for all items

Question 19:

Implant loads:

- Monitor with a BI PCD containing a Type 5 Integrating Indicator. Implants should be quarantined until BI results are known, except in emergency situations.

Not evident

Monitored but not quarantined

Monitored and quarantined

Question 20:

Non-implant loads:

- *Optional* monitoring with a PCD containing either: a BI, a BI and Type 5, a Type 5 integrating indicator, or a Type 6 emulating indicator

Not used immediately

Done for some loads

Done for all loads

Question 21:

Routine steriliser efficacy testing with a BI PCD is done daily (if steriliser run daily):

Sterilisers larger than 60 liter:

- Place BI PCD in first load of items to be sterilised, on bottom shelf of steriliser cart over drain

Not evident

Done sometimes

Done daily when used

Table top sterilisers:

- BI PCD is run with first load of the day and generally placed in centre of load

Use appropriate BI PCD depending on type of steriliser

Question 22:

Steam:

- Daily (each day the steriliser is used)

Used weekly

Used daily

Used every load

SECTION D: Sterilisation and Monitoring (continued)



Use appropriate BI PCD depending on type of steriliser (continued)

Question 23:

Gaseous sterilization (e.g. EO, H2O2):

- BI should be used every load

Not evident

Done but no regular recording

Done with good recording always

Question 24:

Sterilisers larger than 60 liter:

- Use commercially available FDA-cleared BI PCD or AAMI 16-towel pack (M)

Used weekly

Used daily

Used every load

Question 25:

Table top sterilisers:

- BI PCD is a user assembled challenge test pack, which creates the greatest challenge (e.g., BI in peel pouch, BI in wrapped set) and contains items normally processed

Used weekly

Used daily

Used every load

BI test/control and results

Question 26:

Control BI:

- Incubate a positive BI control each day a test vial is incubated and in each auto-reader or incubator. The Control BI needs to be from the same lot number as the Test BI. Record results.

Not evident

Done but no regular recording

Done with good recording always

Question 27:

Test BI:

- Incubate Test BI according to BI Manufacturers' Instructions. Record results.

Not evident

Done but no regular recording

Done with good recording always

SECTION D: STERILISATION AND MONITORING continues >>>

SECTION D: Sterilisation and Monitoring (continued)



Qualification testing

Question 28:

For sterilisation process failures where the cause is not immediately identifiable, and after major steam or steriliser repairs, run 3 empty cycles with a BI PCD followed by 3 empty cycles with a Bowie-Dick test if prevacuum steriliser

Not evident

Done but not as full procedure

Done as in standard

Steriliser maintenance

Question 29:

Steriliser “drain strainers” are inspected daily for debris

Not evident

Occasionally done

Inspected daily

Question 30:

Steriliser external and internal surfaces are routinely cleaned

Not evident

Occasionally cleaned/
some parts cleaned only

Routinely cleaned for external
and internal surfaces

SECTION D: Sterilisation and Monitoring	TOTAL:
	/ 460 points
	MANDATORY ITEMS MARKED:
	/ 320 (16 items)

I completed all questions in this section
especially the **MANDATORY** (highlighted) items.

Yes

No

SECTION E: Sterile Storage and Distribution



Sterile storage

Question 1:

Written policies and procedures are available for storage, handling, rotation, and labelling of sterile packs

- Not available
- Some are available
- All are available

Question 2:

Traffic in the sterile storage area is controlled to limit access to sterile items

- Not controlled
- Controlled but not fully complied by staffs
- Controlled and full compliance

Question 3:

Outside shipping containers and corrugated cartons are not used as containers in sterile storage areas

- Not using appropriate containers
- Partial compliance
- Full compliance

Question 4:

Storage area temperature is generally less than 24°C and relative humidity should not exceed 70%

- Both not complied with
- One of them is complied with
- Both complied with

Question 5:

Sterile items are stored at least 20-25cm (8-10") above the floor, at least 45cm (18") below the ceiling or sprinkler heads, and at least 5cm (2") from outside walls

- Non-compliance
- One of them is compliant
- Full compliance

Question 6:

Shelving and storage carts have a physical barrier between the bottom shelf and the floor

- No barrier
- Partial barrier
- Full physical barrier

SECTION E: STERILE STORAGE AND DISTRIBUTION: Sterile storage continues >>>

SECTION E: Sterile Storage and Distribution (continued)



Sterile storage (continued)

Question 7:

Medical/surgical items, including rigid containers, are not stored next to or under sinks, under exposed water/sewer pipes, or in any location where they may become wet

Non-compliant throughout area

Partial compliance

Full compliance in all areas

Question 8:

Supplies are stored only on designated shelving, counters, and carts (not on windowsills, floors, etc.)

Non-compliant throughout area

Partial compliance

Full compliance in all areas

Question 9:

When stacking container systems, ensure they are firmly seated on one another

Non-compliant throughout area

Partial compliance

Full compliance in all areas

Distribution

Question 10:

Supplies are distributed on a First In First Out (FIFO) basis

Not done

Occasionally done

All the time

Question 11:

Packaging is inspected visually for integrity, and labelling, prior to using items

Both not complied with

One of them is complied with

Both complied with

Question 12:

Transport carts should have a physical barrier between the bottom shelf and the floor
– Reusable covers should be cleaned after each use

Both not complied with

One of them is complied with

Both complied with



SECTION E:
 Sterile Storage and Distribution (continued)

Distribution (continued)

Question 13:

Carts are decontaminated/dried before reused for transporting sterile supplies

- Not done
- Occasionally done
- Done all the time

.....

SECTION E: Sterile Storage and Distribution	TOTAL:
	/ 140 points
	MANDATORY ITEMS MARKED:
	/ 20 (1 item)

I completed all questions in this section especially the **MANDATORY** (highlighted) items.

- Yes
- No

SECTION F: Documentation



Equipment and cycling documentation

Question 1:

Documentation for each mechanical washer is maintained:

- Monitor and verify cleaning processes (e.g. digital readouts, and cycle printouts)

Not done

Some are monitored and documented

All are monitored and documented

Question 2:

Documentation for each steriliser is maintained, and includes results from each load (e.g. monitoring results; steriliser repair records)

Not done

Documentation done occasionally

Documentation done always

Question 3:

For each cycle printout tape:

- Verify cycle start was initiated
- Ensure cycle selected was appropriate for load contents
- Verify correct Time & Temp. was met
- Ensure there were no cycle aborts or warnings

Not evident

Printout tape done occasionally

Printout tape done for each cycle

Question 4:

Record for each cycle:

- Lot number
- Load contents
- Exposure time/temp; *Name/initials of steriliser operator
- Results of BI testing, if applicable
- Results of Bowie-Dick testing, if applicable
- Results of CIs in test packs; reports of non-conclusive or non-responsive CIs found in the load

Not evident

Records done occasionally

Records available for each cycle

Question 5:

An instrument tracking system or other type of computer system is used

No system available

Manual

Electronic

SECTION F: Documentation (continued)



Product recalls

Question 6:

- Policies & Procedures are clear and concise
- Records are maintained
- Lot control labels are used, to include:
Steriliser ID, lot number, sterilisation date,
expiration date, name of pack and initials

No product recall policy

Policy exists but records not
maintained regularly

Policy exists and records
available for all items

Question 7:

Sterilisation Process Failure:

- When cannot immediately identify cause of
failure (e.g. selected incorrect cycle setting),
reprocess the load and recall/reprocess all
items dating back to last load in steriliser with
negative BI results

Not done

Occasionally done

Done all the time

SECTION F: Documentation	TOTAL:
	/ 120 points
	<u>MANDATORY</u> ITEMS MARKED:
	/ 100 (5 items)

I completed all questions in this section
especially the **MANDATORY** (highlighted) items.

Yes

No

SECTION G: Facility Design



Question 1:

All instrumentation reprocessing
is centralized

Not centralized

Centralized

Question 2:

If centralized reprocessing is not possible,
consistent policies and procedures between
locations are in place

No policy

Inadequate policy

Clear comprehensive policy

Question 3:

CSSD department size is appropriately
designed with regard to anticipated volume

Not done

Attempts to do so and has clear
plans to modify and improve

Yes

Question 4:

Decontamination area facilitates proper
workflow and provides adequate space
for necessary equipment

Poor workflow and inadequate space

Inadequate space but good workflow

Proper workflow and adequate space

Question 5:

Decontamination area has space
dedicated to donning and removal
of PPE

No dedicated space

Dedicated space but inadequate

Good adequate dedicated area

Question 6:

Decontamination sink is of adequate size
and has three compartments (for soaking,
cleaning and rinsing)

Inappropriate design

Adequate size but < 3 compartments

Adequate size with 3 compartments
and connectors

Question 7:

Handwashing sinks/hand hygiene facilities
are appropriately located in department

No sinks/hand hygiene facilities

Inappropriately located

Appropriately located

SECTION G: Facility Design (continued)



Question 8:

Emergency eyewash stations (required by OSHA) located within 10 seconds travel time of all chemical usage locations, with a continuous flush for at least 15 minutes (e.g. decontamination area)

Not available
Inappropriately located
Appropriately located

Question 9:

Functional workflow pattern:
– Clear distinction (i.e. physical wall) between dirty and clean

No distinction
Unclear distinction
Clear distinction

Question 10:

Functional workflow pattern:
– Pass-through window available to avoid hallways, and is not propped open

Not available
Available but open
Available and not propped open

Question 11:

Temperature and humidity monitoring controls in decontamination and clean areas

Not available
Only one monitor
Monitors present for both

Question 12:

Temperature and humidity monitoring is recorded daily

Not done
Irregular monitoring
Daily documented monitoring

Question 13:

Appropriate traffic control. Written policy and procedure in place for authorized entry and movement and attire.

No policy
Policy available but not complied with
Policy available with full compliance

Question 14:

Floors and walls are constructed from materials that can withstand frequent cleaning

Poor materials
One of them can withstand frequent cleaning
Both items fulfilled

SECTION G: Facility Design (continued)



Question 15:

Ceilings are flush surfaces and not of materials that are of a particulate or fibre-shedding composition

Did not meet standard

Only fulfill one criteria

Meet both criteria

Question 16:

Doors close freely and do not have thresholds

No doors

Doors have thresholds/don't close freely

Meet both standards

Question 17:

Appropriate positive (clean areas) and negative (soiled areas) pressure ventilation systems in place

No clear pressure differences

Only one area meets standard

Both areas are of appropriate pressure ventilation

Question 18:

Appropriate air-change in decontamination and storage area

Both area does not meet

Only one area meets standard

Both areas are of appropriate meet standard

Question 19:

Lighting adequate for all work areas

Poor lighting throughout

Some have adequate lighting

Adequate lighting in all work areas

**SECTION G:
Facility Design**

TOTAL:

/ 200 points

MANDATORY ITEMS MARKED:

/ 20 (1 item)

I completed all questions in this section especially the **MANDATORY** (highlighted) items.

Yes

No

SECTION H: Considerations



Question 1:

CS supervisory personnel meet minimum recommended qualifications

No recognized qualifications
Local training only
Trained with relevant regional/
international qualifications

Question 2:

CS supervisory personnel maintain competency and participate in departmental continuing education

Not done
Sporadically done
Regular documented programs

Question 3:

CS technicians meet minimum recommended qualifications

No recognized qualifications
Local training only
Has recognized training regionally/
internationally

Question 4:

All new CS personnel receive initial and comprehensive facility and department orientation

Not done
Some received it
All received it and documented

Question 5:

All CS personnel receive a minimum annual training on department policies and procedures All CS personnel demonstrate competency annually

Not done
Not regularly done
All received it and documented

Question 6:

Written policy on personal hygiene

No policy
Inadequate policy
Clear comprehensive policy

SECTION H: CONSIDERATIONS continues >>>

SECTION H: Considerations (continued)



Question 7:

Written policy and adherence to appropriate CS personnel attire

No policy

Inadequate policy/partial compliance

Clear comprehensive policy with full compliance

Question 8:

Written policy and adherence to appropriate PPE in decontamination area

No policy

Inadequate policy/partial compliance

Clear comprehensive policy with full compliance

Question 9:

Written policy and schedule for housekeeping

No policy or schedule

Has policy but no schedule

Has policy and schedule

Question 10:

Written policy and schedule for instrument and sterilizer machine maintenance

No policy

Inadequate policy/partial compliance

Clear comprehensive policy with full compliance

Question 11:

Products used for any/all stages in reprocessing (cleaning, disinfection, sterilization) must be approved by the committee responsible for product selection, by an individual with reprocessing expertise and by an individual with infection prevention and control expertise

No collaboration

Partial collaboration

Full collaboration

**SECTION H:
Considerations**

TOTAL:

/ 120 points

MANDATORY ITEMS MARKED:

/ 20 (1 item)

I completed all questions in this section especially the **MANDATORY** (highlighted) items.

Yes

No

Expert Feedback



Comments:

Impressive elements:

Areas for improvement:

Expert 1 name: _____

Expert 2 name: _____

Date: _____

Summary Scores



Facility name: _____

Country: _____

Section	Item	Score	MANDATORY items
A	Handling, Collection and Transport of Contaminated Instruments	/ 160	/ 100 (5 items)
B	Cleaning and Decontamination Processes	/ 180	/ 80 (4 items)
C	Instrumentation Inspection, Preparation and Packaging	/ 260	/ 100 (5 items)
D	Sterilisation and Monitoring	/ 460	/ 320 (16 items)
E	Sterile Storage and Distribution	/ 140	/ 20 (1 item)
F	Documentation	/ 120	/ 100 (5 items)
G	Facility Design	/ 200	/ 20 (1 item)
H	Considerations	/ 120	/ 20 (1 item)
Grand total		/ 1640	/ 760 (38 items)